



Check Mark Dairy Laboratory

A division of Dairy One
730 Warren Road, Ithaca, NY 14850-1293 - 607-257-1272
Toll Free 1-800-496-3344 - Fax 607-257-6808

Calibration Sample Results

Batch: 1854

Samples Tested: 6/3/21

Multiply all results by 1000.

<u>Sample</u>	<u>DMSCC</u>	<u>10% Range*</u>
Green	166	141 - 191
Brown	427	384 - 470
Pink	744	670 - 818
Clear	1077	969 - 1185

***Range is 15% for the 100k - 200k standard.**