



# **Check Mark Dairy Laboratory**

A division of Dairy One  
730 Warren Road, Ithaca, NY 14850-1293 - 607-257-1272  
Toll Free 1-800-496-3344 - Fax 607-257-6808

## **Calibration Sample Results**

**Batch: 1855**

**Samples Tested: 6/10/21**

**Multiply all results by 1000.**

<b><u>Sample</u></b>	<b><u>DMSCC</u></b>	<b><u>10% Range*</u></b>
<b>Turquoise</b>	<b>170</b>	<b>145 - 196</b>
<b>Orange</b>	<b>430</b>	<b>387 - 473</b>
<b>Yellow</b>	<b>755</b>	<b>680 - 831</b>
<b>Grey</b>	<b>1058</b>	<b>952 - 1164</b>

**\*Range is 15% for the 100k - 200k standard.**