



Check Mark Dairy Laboratory

A division of Dairy One
730 Warren Road, Ithaca, NY 14850-1293 - 607-257-1272
Toll Free 1-800-496-3344 - Fax 607-257-6808

Calibration Sample Results

Batch: 1857

Samples Tested: 6/24/21

Multiply all results by 1000.

<u>Sample</u>	<u>DMSCC</u>	<u>10% Range*</u>
Green	144	122 - 166
Brown	401	361 - 441
Pink	691	622 - 760
Clear	1091	982 - 1200

*Range is 15% for the 100k - 200k standard.