



Check Mark Dairy Laboratory

A division of Dairy One
730 Warren Road, Ithaca, NY 14850-1293 - 607-257-1272
Toll Free 1-800-496-3344 - Fax 607-257-6808

Calibration Sample Results

Batch: 1858

Samples Tested: 7/1/21

Multiply all results by 1000.

<u>Sample</u>	<u>DMSCC</u>	<u>10% Range*</u>
Turquoise	149	127 - 171
Orange	435	392 - 479
Yellow	700	630 - 770
Grey	1016	914 - 1118

***Range is 15% for the 100k - 200k standard.**