



# **Check Mark Dairy Laboratory**

A division of Dairy One  
730 Warren Road, Ithaca, NY 14850-1293 - 607-257-1272  
Toll Free 1-800-496-3344 - Fax 607-257-6808

## **Calibration Sample Results**

**Batch:** 1868

**Samples Tested:** 9/9/21

**Multiply all results by 1000.**

<b><u>Sample</u></b>	<b><u>DMSCC</u></b>	<b><u>10% Range*</u></b>
<b>Blue</b>	<b>165</b>	<b>140 - 190</b>
<b>Purple</b>	<b>393</b>	<b>354 - 432</b>
<b>Red</b>	<b>655</b>	<b>590 - 721</b>
<b>White</b>	<b>956</b>	<b>860 - 1052</b>

**\*Range is 15% for the 100k - 200k standard.**