



Check Mark Dairy Laboratory

A division of Dairy One
730 Warren Road, Ithaca, NY 14850-1293 - 607-257-1272
Toll Free 1-800-496-3344 - Fax 607-257-6808

Calibration Sample Results

Batch: 1873

Samples Tested: 10/14/21

Multiply all results by 1000.

<u>Sample</u>	<u>DMSCC</u>	<u>10% Range*</u>
Turquoise	155	132 - 178
Orange	436	392 - 480
Yellow	723	651 - 795
Grey	1082	974 - 1190

***Range is 15% for the 100k - 200k standard.**