



Check Mark Dairy Laboratory

A division of Dairy One
730 Warren Road, Ithaca, NY 14850-1293 - 607-257-1272
Toll Free 1-800-496-3344 - Fax 607-257-6808

Calibration Sample Results

Batch: 1874

Samples Tested: 10/21/21

Multiply all results by 1000.

<u>Sample</u>	<u>DMSCC</u>	<u>10% Range*</u>
Blue	167	142 - 192
Purple	446	401 - 491
Red	746	671 - 821
White	1050	945 - 1155

***Range is 15% for the 100k - 200k standard.**